



- 110W Wide Operating Temperature range AC-DC Power Supply
- IT (2nd Ed.) and Medical (3rd Ed. MOPP) safety approved, excluding conduction cooling condition
- 2" x 4" standard footprint
- High efficiency up to 93%
- Remote sense
- Remote On/Off (for V1, Negative logic)
- Universal AC input
- Low profile 1U package
- Convection, conduction, and forced-air cooled operation
- Complies with 5000m altitude (at class I IT equipment)
- RoHS compliant
- Less than 0.3W input power at no load remote "off"
- Complies with ErP/Energy star requirement (average efficiency > 87%)



Available now at http://power.murata.com/acdc3d

## **DESCRIPTION**

The MVAD160 series switching power supplies utilize advanced component and circuit technologies to deliver high efficiency. Designed for Medical, Telecom, and Industrial applications to satisfy 1U height design considerations, the MVAD160 Series measures only 2.0" x 4.0" x 1.5". All models offer universal AC input and compliance to worldwide safety and EMC standards.

ORDERING GUIDE								
Model Number	Natural Convection Cooling (@ta = 50°C)	Conduction Cooling (@ta = 85°C)	Forced Air Cooling (@350LFM /ta=50°C)	Main Output (V1)	Aux Output (V2)			
MVAD160-125				12V				
MVAD160-245		110W	160W	24V	5V			
MVAD160-485	110W			48V				
MVAD160-12	TTUW	TIOW	TOUV	12V	no aux output			
MVAD160-24				24V	no remote sense			
MVAD160-48				48V	no V1 remote			

INPUT CHARACTERISTICS						
Parameter	Conditions	Min.	Тур.	Max.	Units	
Input Voltage Operating Range	Single phase	85 <sup>1</sup>	120/230	264	Vac	
Input Voltage Operating Range	DC <sup>2</sup>	200		300	Vdc	
Input Frequency		47	50/60	63	Hz	
Turn-on Input Voltage	Input rising at full load			84		
Turn-off Input Voltage	Input falling at full load			84	Vac	
Input Current	90Vac input, full load			2.1	Α	
Inrush Current	At 264Vac, at 25°C cold start		100		Apk	

OUTPUT CHARACTERISTICS							
Model Number	Main Output Voltage (V1)	Load Current	Load Capacitance	Line, Load, Cross Regulation	Typical Efficiency @230Vac full load		
MVAD160-12x	12V	0 to 13.34A	0 to 1500μF		92%		
MVAD160-24x	24V	0 to 6.67A	0 to 680µF	± 2%	93%		
MVAD160-48x	48V	0 to 3.34A	0 to 330µF		9370		
Model Number	Aux Output Voltage (V2)	Load Current	Load Capacitance	Line, Load, Cross Regulation			
MVAD0160-xx5	5V	0 to 0.1A	0 to 220pF	± 5%			

Main Output Characteristics (all models)						
Parameter	Conditions	Min.	Max.	Units		
Transient Response	50% load step, min.5% load, 1A/µsec slew rate		± 5	%V1		
Settling Time to 1% of Nominal	50% load step, min.5% load, 1A/µsec slew rate		2	msec		
Turn On Delay	After application of input power		1.5	sec		
Output Voltage Rise	Monotonic, <sup>5</sup> 0 to 100% load		200	msec		
Setpoint Accuracy	115Vac, 110W load, 25°C		± 0.5	%V1		
Output Holdup	100% load	10		msec		
Temperature Coefficient			0.02	%/°C		
Ripple Voltage & Noise <sup>3</sup>	min. 5% load		1	%V1		
Remote Sense	Compensates for up to total 400mV of positive and negative lead drops with remote sense connected.		400	mV		

- 1 Refer to power derating curve vs. input voltage. Input test condition on safety approval : 100-240Vac ±10%.
- 2 Applicable for IT Equipment only.
- 3 Ripple and noise are measured with 0.1 µF of ceramic capacitance and 47 µF of electrolytic capacitance on each of the power supply outputs. The output noise requirements apply over a 0 Hz to 20 MHz bandwidth. A short coaxial cable with 50Ω scope termination is used.
- 4 Unless otherwise specified all readings are taken at 115Vac input and 25°C ambient temperature.
- 5 This power supply may exhibit up to 5% turn on overshoot.





Parameter	Conditions	Min.	Typ.	Max.	Units	
Storage Temperature Range		-40 85				
Operating Temperature Dange	Refer to power derating curve vs. operating temperature	-40		85	°C	
Operating Temperature Range	Start up	-40				
Operating Humidity	Non-condensing	10		95	%	
Decrating Altitude	For class I IT equipment deployment	-200		5000	m	
Operating Altitude	Other	-200		2000	111	
MTBF	Telcordia SR-332 M1C3 50°C, 110W	1M			Hours	
Phoek	Operating, IEC60068-2-27, half-sine 5G, 6ms, 3 times per face, 6 faces	1 0 / Complies				
Shock	Non-operating, IEC60068-2-27, half-sine, 30G, 18ms, 3 times per face, 6 faces	Linmiles				
Florida	Operating, IEC60068-2-6, 1.0G, 10-150Hz, 10minutes per axis, on all 3 axes					
/ibration	Non-operating, IEC60068-2-6, 2.0G, 10-150Hz, 10minutes per axis, on all 3 axes					
Safety	IEC60950-1:2005 2nd Ed.; Am1:2009 EN60950-1:2006; A11:2009; A1:2010; A12:2011 UL60950-1, 2nd Ed., 2011-12-19 & CSA C22.2 No. 60950-1-07, 2nd Ed., 2011-12 IEC60601-1:2005 3rd Ed. + CORR.1(2006) + CORR.2(2007) ANSI/AAMI ES60601-1 (2005+C1:09 + A2:10), CSA-C22.2 No. 60601-1(2008), MOPP CE Marking per LVD					
Narranty Parranty	2 years <sup>6</sup>					
Outside Dimensions	2.0" x 4.0" x 1.5" (50.8mm x 101.6mm x 38.1mm)					
Veight	0.42lbs (190g) typical					

PROTECTION CHARACTERISTICS					
Parameter	Conditions	Min.	Тур.	Max.	Units
Overvoltage Protection	V1; Latching (50% load)	110		140	
	V2; Latching (50% load)	110		170	%
Overcurrent Protection	V1; Hiccup mode; Auto Recovery	110		170	
Overcurrent Protection	V2; Hiccup mode; Auto Recovery	0.11	0.4		Α
Overtemperature Protection	Auto recovery		Complies		

ISOLATION CHARACTERISTICS						
Parameter	Conditions	Min.	Typ.	Max.	Units	
	Primary to Earth Ground (1x MOPP)	1500				
Isolation	Primary to Secondary (2x MOPP) <sup>7</sup>	4000			Vac	
	Secondary to Earth Ground	500			Vdc	
Lankana Cumant (undan naman) aanditiana)	264Vac, 60Hz, 25°C			350	μА	
Leakage Current (under normal conditions)	240Vac, 60Hz, 25°C			300	μA	
Touch Current	264Vac, 60Hz, 25°C			100	μА	

Standard	Compliance
IEC/EN 61000-3-2	Class D, Class C
IEC/EN 61000-3-3	Complies
EN 55022	Class B, (Class A as Class II equipment deployment)
FCC Part 15	Class B, (Class A as Class II equipment deployment)
IEC/EN 61000-4-2	Level 4, Criterion A
IEC/EN 61000-4-3	Level 2, Criterion A
IEC/EN 61000-4-4	Level 3, Criterion A
IEC/EN 61000-4-5	Level 4, Criterion A
IEC/EN 61000-4-6	Level 2, Criterion A
IEC/EN 61000-4-8	Level 2, Criterion A
IEC/EN 61000-4-11	Level 3, Criterion B
	IEC/EN 61000-3-2 IEC/EN 61000-3-3 EN 55022 FCC Part 15 IEC/EN 61000-4-2 IEC/EN 61000-4-3 IEC/EN 61000-4-4 IEC/EN 61000-4-5 IEC/EN 61000-4-6 IEC/EN 61000-4-8

6 At Ta<50°C.

7 Class I equipment deployment.

ecifications are subject to change without notice. It is responsibility of each customer to thoroughly test each product and part number under their unique parameters and environments to ensure a product will work properly and reliab



CONTROL SIGNAL	
Parameter	Conditions
V1 Remote	This signal must be pulled low state (maximum of 0.2Vdc, sink current >15mA) to V2 return terminal to turn on the main V1 output. The power supply shall include an internal pull up resistor to an internal 12Vdc to disable the main V1 output if the signal is left open. The power supply shall not pull this signal down if there is no AC input. The open circuit voltage present when in the "open" state shall not exceed 15Vdc.
	The AUX output is independent of the V1 remote signal.

# **EMI CONSIDERATIONS**


For optimum EMI performance, the power supply should be mounted to a metal plate grounded to all 4 mounting holes of the power supply. To comply with safety standards, this plate must be properly grounded to protective earth (see mechanical dimension notes). Pre-compliance testing has shown that a standalone power supply complies with EN55022 class A radiated emissions. Radiated emission results vary with system enclosure and cable routing paths.


#### **SAFETY CONSIDERATIONS**




- 1. This power supply is a component level power supply intended for use in Class I or Class II applications. Secondary ground traces need to be suitably isolated from primary ground traces when used in class II applications.
- 2. When the power supply is used in Class II equipment, all ground traces and components connected to the primary side are considered primary for spacing and insulation considerations.


# PERFORMANCE DATA













# **MECHANICAL DIMENSIONS**





All dimensions in inches (mm). Tolerance is +/-0.02" (0.5)

Mounting holes H3 and H4 should be earth-grounded for EMI purpose.

Mounting holes H4 is earth ground connection

This power supply requires mounting on standoffs minimum 0.197" (5.0) in height for forced air cooling.


Dimensions: 2.0" x 4.0" x 1.5" (50.8mm x 101.6mm x 38.1mm)

INPUT/OUT	PUT CONNECTOR AND SIGNAL SPECIFI	CATION AND MATING CONNECTORS		
Pin	Description	Mating Housing	Crimp terminal/pins	Outline
Input Cor	nnector CN1: TE Connectivity 5-1376382-1			A MACK MARK
1	Line (V-)	TE Connectivity 1376388-1	TE Connectivity 1376348-1	
2	Neutral (V+)			
	Spade Connector: #250			
GND	Earth Ground			^ <u></u>
Outp	ut Connector CN2: Molex 39-28-8040			( <u>na</u> )
1, 2	V1	Molex 39-01-2040	Molex 39-00-0038	34
3, 4	V1 Return			
Outp	ut Connector CN3: Molex 87438-0643			Circuit #1
	for models with auxiliary output			PITCH → 1.50 ←
1	V2			
2	V2 Return	Molex 87439-0600	Molex 87421-0000	
3	V1 Remote	WIDIEX 07439-0000	WILLEY 01451-0000	<del>                                    </del>
4	V1 -Sense			
5	N.C			
6	V1 +Sense			



# THERMAL CONSIDERATIONS

**Conduction Cooling Image (and Preferred Interface Materials)** 



#### Aluminum Baseplate:

Thickness 2.5 to 3mm

Maximum operating temperature +85°C (measurement point: center on the back side of baseplate)

## Preferred mounting on standoffs: 3.5mm

#### Isolation sheet or Isolation tape

Preferred materials: thickness 50 to 130um, 94V-0

#### Thermal Sheet or Thermal Gel:

Preferred Materials:

Type: Sarcon 300G-ae (thickness 3mm, thermal conductivity 1.3W/m·K) Manufacturer: Fuji Polymer

Type: GEL30 (thermal conductivity 3.5W/m·K)

Manufacturer: Parker

### **Forced Air Cooling**

System thermal management is critical to the performance and reliability of the MVAC series power supplies. Performance derating curves are provided which can be used as a guideline for what can be achieved in a system configuration with controlled airflow at various input voltage conditions.

The air flow curves are generated using an AMCA 210-99 and ASHRAE 51-1999 compliant wind tunnel with heated inlet air and a controlled CFM providing a duct test section having a calculated average LFM. A correlation between the test setup and the actual system environment is paramount to understanding what can be achieved in an actual system. In a power supply of this density, cooling air moving both through the unit as well as around the unit strongly influences local temperatures. The wind tunnel test setup was constructed to produce a flow with a slight back pressure to induce both flow conditions by providing a small gap between the power supply and duct walls of 0.5" (13mm). The optimal and characterized airflow direction is from the input connector to the output connector (see diagram below). The P-Q flow curve for this test setup is also shown below.

